



**Sodium chloride calibration
solution for refractometer
verification**

Abbreviation	% Sodium chloride (w/w)
K10	10,0 ± 0,10
K15	15,0 ± 0,15
K20	20,0 ± 0,20